

CERTIFICATE OF COMPLIANCE

Certificate Number 2019-10-16-E492083
Report Reference E492083-D1000-1/A0/C0-ULCB
Issue Date 2019-10-16
Issued to: MGF S.r.l.
Applicant Company: Via Pascoli 15
Abbiategrosso (MI) 20081 ITALY
Listed Company: Same as Applicant

**This is to certify that
representative samples of**

Air compressor for dental applications
CS 50/25 PRIME X, CS 50/25 PRIME M, CS 50/25 PRIME S,
50/25 PRIME X, 50/25 PRIME M, 50/25 PRIME S, CS 30/15
PRIME M, CS 30/15 PRIME S, 30/15 PRIME M, 30/15 PRIME S,
50/15 PRIME M, 50/15 PRIME S, MINI BOX 24/7 PRIME M,
MINI BOX 24/7 PRIME S, MINI BOX EVO 24/7 PRIME M, MINI
BOX EVO 24/7 PRIME S

Have been investigated by UL in accordance with the
Standard(s) indicated on this Certificate.

Standard(s) for Safety: ANSI/AAMI ES60601-1: A1:2012, C1:2009/(R)2012 and
A2:2010/(R)2012, CSA CAN/CSA-C22.2 NO. 60601-1:14, IEC
60601-1 Edition 3.1 (2012)

Additional Standards: IEC 60601-1-6:2010 (Third Edition) + A1:2013

Additional Information: See the UL Online Certifications Directory at
www.ul.com/database for additional information.


Only those products bearing the UL Certification Mark should be considered as being covered by UL's
Certification and Follow-Up Service.

Look for the UL Certification Mark on the product.

This is to certify that representative samples of the product as specified on this certificate were tested
according to the current UL requirements.



Bruce Mahrenholz, Assistant Chief Engineer, Global Inspection and Field Services, UL LLC



Helena Y. Wolf, Director, Global Market Access Operations, UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL
Customer Service Representative www.ul.com/contactus

